UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/830,179	04/22/2004	Stuart L. Axelson JR.	OSTEONICS 3.0-488	7711
	7590 03/03/200 /ID, LITTENBERG,	EXAMINER		
KRUMHOLZ &	& MENTLIK		ARAJ, MICHAEL J	
600 SOUTH AVENUE WEST WESTFIELD, NJ 07090			ART UNIT	PAPER NUMBER
			3733	
			MAIL DATE	DELIVERY MODE
			03/03/2008	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

	Application No.	Applicant(s)				
Office Action Comments	10/830,179	AXELSON ET AL.				
Office Action Summary	Examiner	Art Unit				
	MICHAEL J. ARAJ	3733				
The MAILING DATE of this communication appears on the cover sheet with the correspondence address Period for Reply						
A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.  - Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.  - If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.  - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).						
Status						
1) Responsive to communication(s) filed on <u>06 De</u>	ecember 2007					
· <u> </u>	· · · · · · · · · · · · · · · · · · ·					
<i>,</i>	<del>/ -</del>					
•	closed in accordance with the practice under <i>Ex parte Quayle</i> , 1935 C.D. 11, 453 O.G. 213.					
closed in accordance with the practice under L	x parte Quayle, 1900 C.D. 11, 40	0.0.210.				
Disposition of Claims						
4)⊠ Claim(s) <u>1,2,6-9,11-16,19,20,25-27</u> is/are pend	Claim(s) <u>1,2,6-9,11-16,19,20,25-27</u> is/are pending in the application.					
4a) Of the above claim(s) is/are withdraw	4a) Of the above claim(s) is/are withdrawn from consideration.					
5) Claim(s) 9,11,12,15,16,19,20 and 27 is/are allowed.						
6)⊠ Claim(s) <u>1,2,6-8,13,14,25 and 26</u> is/are rejected.						
7) Claim(s) is/are objected to.						
•	election requirement					
8) Claim(s) are subject to restriction and/or election requirement.						
Application Papers						
9)☐ The specification is objected to by the Examiner.						
10) The drawing(s) filed on is/are: a) accepted or b) objected to by the Examiner.						
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).						
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).						
11)☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.						
Priority under 35 U.S.C. § 119						
a) ☐ All b) ☐ Some * c) ☐ None of:	12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).					
<i>,</i> — <i>,</i> — <i>,</i> —						
	1. Certified copies of the priority documents have been received.					
2. Certified copies of the priority documents have been received in Application No						
3. Copies of the certified copies of the priority documents have been received in this National Stage						
application from the International Bureau (PCT Rule 17.2(a)).						
* See the attached detailed Office action for a list of the certified copies not received.						
Attachment(s)						
1) Notice of References Cited (PTO-892)  4) Interview Summary (PTO-413)						
2) Notice of Draftsperson's Patent Drawing Review (PTO-948)  Paper No(s)/Mail Date  3) Information Disclosure Statement(s) (PTO/SB/08)  Notice of Informal Patent Application						
3) Information Disclosure Statement(s) (PTO/SB/08)  Paper No(s)/Mail Date  5) Notice of Informal Patent Application  6) Other:						
<del></del>	· <del></del>					

#### **DETAILED ACTION**

The final mailed on September 6, 2007 has been withdrawn due to the following action.

## Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1, 6-8, 13, 14, 25 and 26 are rejected under 35 U.S.C. 102(b) as being anticipated by Zarnowski et al. (U.S. Patent No. 4,892,093).

Zarnowski et al. disclose a cutting block comprising a body (20) having a first generally planar contacting surface, a second surface spaced from said first bone contacting surface along an axis perpendicular to said generally planar bone contacting surface, said body having a perimeter surface extending between said first and second surfaces wherein at least a medial and a lateral perimeter surface of the body are curved along a plane parallel to the planar bone contacting surface and along a plane perpendicular to the bone contacting surface (refer to Figure 1 below). The body is used to resect a distal femur and includes a plurality of slots for guiding a saw blade, wherein the body includes slots for forming an anterior cut, a posterior cut, an anterior chamber cut and posterior chamfer cut in the distal femur and wherein the body is made of a plurality of pieces connected by pins and space to form said slots (see Fig. 6). The pins are located at adjacent ends of said slots used cuts located adjacent a center of

Art Unit: 3733

said slots. Said slots for forming said posterior cuts intersect said perimeter surface to form an open end. The pins extend from a bottom to a top surface of said slot, wher said pin adjacent a central position of said slot opened end and having a rounded outer surface for engaging the saw blade. The cutting block perimeter surface is shaped to conform to a distal femur. The cutting block is less than or equal to a medial-lateral dimension of an outer perimeter of the distal femur and a height of said cutting block is less that or equal to an anterior-posterior dimension of the distal femur. The distal femur discussed can be a distal femur of an elephant.




Figure 1

# Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claim 2 is rejected under 35 U.S.C. 103(a) as being unpatentable over Zarnowski et al. (U.S. Patent No. 4,892,093) in view of Buttermilch et al. (U.S. Patent No. 6,585,771).

Zarnowski et al. disclose the claimed invention except for the perimeter surface being polised. Buttermilch et al. disclose the use of polished surfaces when dealing with bone to prevent additional damage to the bone. It would have been obvious to one skilled in the art at the time the invention was made to have created the device of Zarnowski et al. with polished surfaces in view of Buttermilch et al., in order to have a device that can be used to cut the bone and eliminate further unwanted damage.

#### Allowable Subject Matter

Claims 9, 11, 12, 15 16, 19, 20 and 27 allowed.

### Response to Arguments

Applicant's arguments with respect to claim 1 have been considered but are moot in view of the new ground(s) of rejection.

Application/Control Number: 10/830,179 Page 5

Art Unit: 3733

It has also been noted that new claim 25 does not include all the limitations of the base claim and any intervening claims, which in this case would be claims 1, 3, 4 and 5.

Claim 25 only includes the limitations from the base claim 1 and dependent claim 5.

#### Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to MICHAEL J. ARAJ whose telephone number is (571)272-5963. The examiner can normally be reached on M-F 8am-5pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on 571-272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/830,179 Page 6

Art Unit: 3733

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Michael J Araj/ Examiner, Art Unit 3733

/Eduardo C. Robert/ Supervisory Patent Examiner, Art Unit 3733